



“ion pulse” simulated with Stanford DG535 and series of fixed attenuators.

Nim output show is at minimum width setting. Minimum width gives maximum pulse pair resolution which in this case is approximately 5 nS.

Screen shot taken with Tektronix TDS7404, 20GS/S 4GHz DPO